

## Activity Assay for Mouse PAI-1

For Research Use Only

---

### INTRODUCTION

---

Plasminogen Activator Inhibitor-1 (PAI-1) is a glycoprotein and member of the serine proteinase inhibitor (serpin) superfamily. PAI-1 is the primary inhibitor of tissue-type plasminogen activator (tPA) and the urokinase-type plasminogen activator (uPA). This inhibition exhibits antiproteolytic properties that can lead to myocardial infarction and thromboembolic disease with elevated levels of PAI-1. Additionally, PAI-1 is thought to play a role in the function of tissue remodeling and tumor metastasis.

---

### PRINCIPLES OF PROCEDURE

---

This is an ELISA (Enzyme-Linked Immunosorbent Assay) for the quantitative analysis of PAI-1 levels in biological fluid. This test kit operates on the basis of sandwich ELISA where active PAI-1 forms complexes with uPA and is quantified with the use of an HRP labeled secondary antibody.

The functional or active PAI-1 binds to the uPA coated on the well of the microtiter plate. The latent and complexed forms of PAI-1 will not bind and are discarded at a later washing step. Next, PAI-1 primary antibody is added to the wells, binding to the captured PAI-1 on the microtiter plate. HRP conjugated secondary antibody is then added for detection of the active PAI-1. Quantitative test results are obtained by the measure and comparison of the sample and standard absorbance readings.

---

### MATERIALS PROVIDED

---

Component	Volume	Storage
Anti-PAI-1 Coated Plate		4°C
Mouse PAI-1 Standard (50 ng/mL)	1 vial	4°C
10x Wash Buffer	50 mL	4°C
Anti-PAI-1 Primary Antibody	1 vial	4°C
TMB Substrate	10 mL	4°C
HRP Conjugated Secondary Antibody	1 vial	4°C

---

### MATERIALS NEEDED BUT NOT PROVIDED

---

- 1 N H<sub>2</sub>SO<sub>4</sub>
- TBS Buffer (see Reagent Preparation)
- Blocking Buffer (see Reagent Preparation)
- DI Water
- Microplate reader with 450 nm filter
- Microplate shaker with uniform horizontal circular movement up to 300 rpm
- Precision pipettes that range from 10 µL-1000 µL and disposable tips

---

### STORAGE

---

- Store the kit and all of its components at 4°C before use.
- If not using the entire plate at once, prepare only the appropriate amount of Primary Antibody and PAI-1 Standard. The remaining stock solutions should be frozen and stored at -70°C.

Primary Antibody should be used within two weeks. All other components should remain refrigerated.

3. Store unused portions of the microplate in a pouch with a desiccant at 4°C.

---

## WARNINGS AND PRECAUTIONS

---

1. This kit is designed to work properly as provided and instructed. Additions, deletions or substitutions to the procedure or reagents are not recommended, as they may be detrimental to the assay.
2. Exercise universal precautions during the performance or handling of this kit or any component contained therein.

---

## PROCEDURAL NOTES

---

1. This assay should be run at room temperature.
2. Use aseptic technique when opening and dispensing reagents.
3. To minimize error due to handling, wipe the exterior bottom of the microplate wells with a lint-free paper towel.

---

## SAMPLE COLLECTION, STORAGE, AND PREPARATION

---

Samples should be collected using trisodium citrate, acidified citrate or Stabilyte™ (DiaPharma) collection media. Collection should be in accordance with the collection vials manufacturers instructions or in a 1:10 ratio of collection media to blood.

Immediately, upon collection of blood, the samples should be centrifuged at 3000 x g. This should ensure the removal of platelets as they can release PAI-1 that, in turn, complexes with uPA. The plasma can be transferred to a clean plastic tube and stored frozen for up to one month. Samples are stable for approximately 24 hours when stored at 4°C with the Sabilyte™ media or one month if stored at -20°C.

**Note:** Detergents such as Triton X cause interference with the assay. If using detergent extracted samples, it is necessary to dialyze the samples overnight to remove the detergent.

---

## REAGENT PREPARATION

---

The following solutions should be prepared fresh before starting the assay.

1. **TBS Buffer:** 0.10 M TRIS, 0.15 M NaCl, pH 7.4
2. **Blocking Buffer:** 3% BSA in TBS Buffer.
3. **10x Wash Buffer:** Dilute to 1x prior to use. Do this by combining 1 part of 10x Wash Buffer to 9 parts of DI Water relative to the amount required for the assay, either in whole or in part.

---

## ASSAY PROCEDURE

---

1. Prepare standards as indicated in the provided dilution table.  
**Note:** The standards should be applied to the plate immediately upon preparation.
2. Add 100 µL of standards or unknowns to each well omitting wells for later use. See **Scheme I** for a suggested template design.
3. Shake the plate at 300 rpm on a plate shaker for 30 minutes.
4. Wash wells according to the following wash procedure:
  - a. Remove contents of the plate by inversion into an appropriate disposal device.
  - b. Tap out remaining contents of the plate onto a lint free paper towel.
  - c. Add 300 µL of Wash Buffer to each well.
  - d. Let stand for 2-3 minutes.
  - e. Repeat procedure 2 more times and proceed to step “f”.
  - f. Remove contents of the plate by inversion into an appropriate disposal device.
  - g. Tap out the remaining contents of the plate onto a lint free paper towel and proceed to step 6.

**Note:** The decanted wells should be void of visible moisture before proceeding. If moisture is still visible then follow step “g” until satisfactory results are obtained.

5. Make a working concentration of Primary Antibody by combining with 3% BSA Blocking Buffer. See vial label for the specific volume.
6. Add 100  $\mu$ L of diluted Primary Antibody to each well.
7. Shake plate at 300 rpm on the plate shaker for 30 minutes.
8. Wash wells according to step 5 located above in this section.
9. Make a working concentration of Secondary Antibody by combining with 3% BSA Blocking Buffer. See vial label for the specific volume.
10. Add 100  $\mu$ L of the working concentration Secondary Antibody solution to each well.
11. Shake plate at 300 rpm on the plate shaker for 30 minutes.
12. Wash wells according to step 5 located above in this section.
13. Add 100  $\mu$ L of TMB Substrate to each well and incubate for 10 minutes. If accounting for substrate background, use 2 wells as blanks with only substrate in the wells (150  $\mu$ L/well).
14. Stop the reaction with 50  $\mu$ L per well of 1 N  $H_2SO_4$  and read plate at 450 nm.


#### Scheme I:

	1	2	3	4	5	6	7	8	9	10	11	12
A	S0	S1	S2	S3	S4	S5	S6	S7	S8	S9	S10	U1
B	S0	S1	S2	S3	S4	S5	S6	S7	S8	S9	S10	U1
C	U2	U3	U4	U5	U6	U7	U8	U9	U10	U11	U12	U13
D	U2	U3	U4	U5	U6	U7	U8	U9	U10	U11	U12	U13
E	U14	U15	U16	U17	U18	U19	U20	U21	U22	U23	U24	U25
F	U14	U15	U16	U17	U18	U19	U20	U21	U22	U23	U24	U25
G	U26	U27	U28	U29	U30	U31	U32	U33	U34	U35	U36	BLK
H	U26	U27	U28	U29	U30	U31	U32	U33	U34	U35	U36	BLK

#### CALCULATIONS

1. Subtract the average O.D. value of the blank wells (BLK) from all other pairs of wells. Some microplate readers can be programmed to do these subtractions automatically when reading the plate. Consult your instrument manual.
2. Average the O.D. values for each pair of duplicate wells.
3. Plot a standard curve using the average O.D. value for each standard value versus the concentration of standard.
4. Determine the concentration of each unknown by interpolation from the standard curve.

#### Typical Standard Curve:



---

**PERFORMANCE CHARACTERISTICS**

---

Assay Range: 0.05-50 ng/mL

Samples with uPA levels higher than 50 ng/mL should be diluted in similar media devoid of active PAI-1.

---

**REFERENCES**

---

1. Declerck, P., *et al.*, (1995) *Thromb Haemost.* Nov:74(5):1305-1309
2. Eitzman, D., *et al.*, (2000) *Blood* 95(2):577-580
3. Kawasaki, T., *et al.* (2000) *Blood* 96(1):153-160
4. Declerck, P., *et al.*, (1988) *Blood* 71(1):220-225
5. Schafer, K., *et al.*, (2001) *FASEB* 15:1840-1842
6. Samad, F., *et al.*, (1996) *Proc Natl Acad Sci USA* 96(12):6902-6907

---

**DISCLAIMER**

---

This information is believed to be correct but does not purport to be all-inclusive and shall be used only as a guide. Oxford Biomedical Research, Inc. shall not be held liable for any damage resulting from handling or from contact with the above product. See catalog for additional terms and conditions of sale.

---

**ORDERING INFORMATION**

---

For additional kits or a complete catalog please call 1-800-692-4633.

---

**TECHNICAL SUPPORT**

---

If you need technical information or assistance with assay procedures, call our Technical Support Department at 1-800-692-4633 or 1-248-852-8815. Our staff will be happy to answer your questions about this or any other product in the Oxford Biomedical line.

---

**GUARANTEE AND LIMITATION OF REMEDY**

---

Oxford Biomedical Research, Inc. makes no guarantee of any kind, expressed or implied, which extends beyond the description of the material in this ELISA kit, except that these materials and this kit will meet our specifications at the time of delivery. Buyer's remedy and Oxford Biomedical Research, Inc.'s sole liability hereunder is limited to, at Oxford Biomedical Research, Inc.'s option, refund of the purchase price of, or the replacement of, material that does not meet our specification. By acceptance of our products, Buyer indemnifies and holds Oxford Biomedical Research, Inc. harmless against, assumes all liability for the consequence of its use or misuse by the Buyer, its employees, or others. Said refund or replacement is conditioned of Buyer notifying Oxford Biomedical Research, Inc. within (30) days of the receipt of product. Failure of Buyer to give said notice within said thirty (30) days shall constitute a waiver by the Buyer of all claims hereunder with respect to said material(s).

Oxford Biomedical Research, Inc.  
P.O. Box 522  
Oxford, MI 48371 U.S.A.

Orders: 800-692-4633  
Technical Service: 248-852-8815  
Fax: 248-852-4466  
E-mail: [info@oxfordbiomed.com](mailto:info@oxfordbiomed.com)

Made in the U.S.A.